Display Report

Analysis Info

Analysis Name	D:\Data\2023\May\15\VAA_448.d	Acquisition Date	15.05.2023 10:09:16
Method	tune_low.m	Operator	Bruker Customer
Sample Name		Instrument / Ser#	micrOTOF 10223
Comment			

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.6 Bar
Focus	Not active			Set Dry Heater	200 °C
Scan Begin	100 m/z	Set Capillary	4500 V	Set Dry Gas	6.0 l/min
Scan End	1000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source



